

# Lin-Zhi International, Inc.



Dedicated to the development of  
innovative and quality products for  
clinical diagnostics.

January 11, 2019

Dear End User,

*Lin-Zhi International Inc.* (LZI) apologizes for the inconvenience caused by the recall on the following products.

<b>Reference Number</b>	<b>Description</b>
0220	Ethyl Alcohol Assay Kit, small test kit
0221	Ethyl Alcohol Assay Kit, large test kit

Please be aware that through internal testing, LZI has discovered a potential issue where the shelf-life of the products may be reduced due to degradation of the assay.

The recall affects the products with lot number 1810026 and 1810025, which you had received between October 17, 2018 and January 08, 2019. LZI has determined these are the only products affected and there is no issue with previous lots.

If you received the products from these lots and still have remaining product left, please properly dispose of it and notify your distributor once such action has been completed in order to receive a replacement product or other compensation method indicated by your distributor. Thank you for your patience and understanding.

Sincerely,

Bernice Lin  
VP of Operations  
***Lin-Zhi International, Inc.***  
2945 Oakmead Village Court,  
Santa Clara, CA 95051 USA

T: (480) 970-8811  
F: (408) 970-9030  
E: [regulatory@lin-zhi.com](mailto:regulatory@lin-zhi.com)  
[www.lin-zhi.com](http://www.lin-zhi.com)

***Lin-Zhi International, Inc.***  
2945 Oakmead Village Court  
Santa Clara, CA 95051  
USA

ISO:13485 • CE • CMDCAS • TK • FDA

Tel: (408) 970-8811  
Fax: (408) 970-9030  
E-mail: [customerservice@lin-zhi.com](mailto:customerservice@lin-zhi.com)  
Webpage: [www.lin-zhi.com](http://www.lin-zhi.com)