

LZI Propoxyphene Enzyme Immunoassay

IVD For In Vitro Diagnostic Use Only



REF 0120 (100/37.5 mL R₁/R₂ Kit)
0121 (1000/375 mL R₁/R₂ Kit)



Lin-Zhi International, Inc.

Intended Use

The Lin-Zhi International, Inc. (LZI) Propoxyphene (PPX) Enzyme Immunoassay is intended for the qualitative and semi-quantitative determination of propoxyphene in human urine at a cutoff value of 300 ng/mL. The assay is designed for prescription use with a number of automated clinical chemistry analyzers.

The assay provides only a preliminary analytical result. A more specific alternative analytical chemistry method must be used in order to obtain a confirmed analytical result. Gas or Liquid Chromatography/Mass Spectrometry (GC/MS or LC/MS) are the preferred confirmatory methods (1, 2). Clinical consideration and professional judgment should be exercised with any drug of abuse test result, particularly when the preliminary test result is positive.

Summary and Explanation of Test

Propoxyphene, or dextropropoxyphene, the active principle compound in the prescription drug Darvon, was one of the most prescribed pain relievers for mild to moderate discomfort. In 2010, the FDA requested that propoxyphene be voluntarily pulled from the market. Since then, this drug has been taken off the market in both Europe and the United States due to concerns of fatal overdoses and arrhythmia (3).

Chemically, it resembles methadone, and is used as an alternative for detoxification and maintenance of narcotic dependence (4). However, propoxyphene and its metabolites are cardiotoxic; the likelihood of overdose is higher than methadone. Overdosage can result in convulsions, respiratory depression, cardiac arrhythmia, hypertension, pulmonary edema, circulatory collapse and death (5-7). Propoxyphene is primarily metabolized to norpropoxyphene via N-demethylation. Norpropoxyphene is a biologically active analgesic, although it is less potent than propoxyphene. Further demethylation and dehydration leads to the cyclic dinorpropoxyphene metabolite. Other polar metabolites are formed by aryl hydroxylation, ester hydrolysis, and glucuronide conjugation (8). The rate of clearance varies from person to person; however, approximately 34 % of the administered dosage is eliminated within 20 hours (8, 9), and previous studies have reported that up to 75% is secreted in the urine over a seven day period (10). Detection of propoxyphene or its metabolites in urine indicates use of propoxyphene.

Assay Principle

The Propoxyphene assay is a homogeneous enzyme immunoassay ready-to-use liquid reagent. The assay is based on competition between drug in the sample and drug labeled with the enzyme glucose-6-phosphate dehydrogenase (G6PDH) for a fixed amount of antibody in the reagent (11). Enzyme activity decreases upon binding to the antibody, and the drug concentration in the sample is measured in terms of enzyme activity. In the absence of drug in the sample, propoxyphene-labeled G6PDH conjugate is bound to antibody, and the enzyme activity is inhibited. On the other hand, when drug is present in the sample, antibody binds to the free drug; the unbound propoxyphene-labeled G6PDH then exhibits its maximal enzyme activity. Active enzyme converts nicotinamide adenine dinucleotide (NAD) to NADH, resulting in an absorbance change that can be measured spectrophotometrically at a 340 nm primary wavelength.

Reagents Provided

Antibody/Substrate Reagent (R₁): Contains mouse monoclonal anti-propoxyphene antibody, glucose-6-phosphate (G6P), nicotinamide adenine dinucleotide (NAD), stabilizers, and sodium azide (0.09 %) as a preservative.

Enzyme-drug Conjugate Reagent (R₂): Contains glucose-6-phosphate dehydrogenase (G6PDH) labeled with propoxyphene in buffer with sodium azide (0.09 %) as a preservative.

Calibrators and Controls*

*Calibrators and controls are sold separately and contain negative human urine with sodium azide as a preservative.

PROPOXYPHENE Calibrators	REF
Negative Calibrator	0001
Low Calibrator: Contains 150 ng/mL propoxyphene	0122
Cutoff Calibrator: Contains 300 ng/mL propoxyphene	0123
Intermediate Calibrator: Contains 600 ng/mL propoxyphene	0124
High Calibrator: Contains 1000 ng/mL propoxyphene	0125
PROPOXYPHENE Controls	REF
Level 1 Control: Contains 225 ng/mL propoxyphene	0127
Level 2 Control: Contains 375 ng/mL propoxyphene	0128

Precautions and Warning

- This test is for in vitro diagnostic use only. Harmful if swallowed.
- Reagent contains sodium azide as a preservative, which may form explosive compounds in metal drain lines. When disposing such reagents or waste, always flush with a large volume of water to prevent azide build-up. See National Institute for Occupational Safety and Health Bulletin: Explosive Azide Hazards (12).
- Do not use the reagents beyond their expiration dates.
- For USA: Caution: Federal law restricts this device to sale by or on the order of a physician.

Reagent Preparation and Storage

The reagents are ready-to-use. No reagent preparation is required. All assay components should be refrigerated at 2-8°C when not in use.

Specimen Collection and Handling

Urine samples may be collected in plastic or glass containers. Some plastics may absorb drugs. Use of plastics such as polyethylene is recommended (13). Use fresh urine specimens for the test. If a sample cannot be analyzed immediately, it may be refrigerated at 2-8°C for up to seven days (14, 15). For longer storage, keep sample frozen at -20°C and then thaw before use. Studies have shown propoxyphene analytes in urine are stable at -20°C for up to 6 months (15). Samples should be at room temperature (18-25°C) for testing. Samples with high turbidity should be centrifuged before analysis. Fresh and properly stored urine samples are generally within this range. Adulteration may cause erroneous results. If sample adulteration is suspected, obtain a new sample and forward both samples to a laboratory for testing. Handle all urine specimens as if they are potentially infectious.

Instrument

Clinical chemistry analyzers capable of maintaining a constant temperature, pipetting samples, mixing reagents, measuring enzyme rates at a 340 nm primary wavelength and timing the reaction accurately can be used to perform this homogeneous immunoassay.

Performance characteristics presented in this package insert have been validated on the Hitachi 717. If other instruments are used, performance needs to be validated by the laboratory (16, 17).

Assay Procedure

Analyzers with the specifications indicated above are suitable for performing this homogeneous enzyme immunoassay. Refer to the specific parameters used for each analyzer before performing the assay. Typical assay parameters used for the Hitachi 717 analyzer include a 20 µL sample, 200 µL of antibody reagent (R₁), and 75 µL of enzyme conjugate reagent (R₂) in 37°C incubation temperature, 30-35 reading frames, and a 340 nm primary wavelength. For qualitative analysis, use the 300 ng/mL as the cutoff calibrator. For semi-quantitative analysis, use all five calibrators. Recalibration should be performed after reagent bottle change or if there is a change in calibrators or reagent lot. Two levels of controls are also available for monitoring of the cutoff level: 225 ng/mL and 375 ng/mL.

Calibration and Quality Control

Good laboratory practices recommend the use of at least two levels of control specimens (one positive and one negative control near the cutoff) to ensure proper assay performance. Controls should be run with each new calibration and after specific maintenance or troubleshooting procedures as detailed in the instrument system manual. Each laboratory should establish its own control frequency. If any trends or sudden change in control value are observed, review all operating parameters, or contact LZI technical support for further assistance. Laboratories should comply with all federal, state, and local laws, as well as all guidelines and regulations.

Results

Note: A preliminary positive test result does not necessarily mean a person took illegal drugs and a negative test result does not necessarily mean a person did not take illegal drugs. There are a number of factors that influence the reliability of drug tests.

Qualitative: The cutoff calibrator, which contains 300 ng/mL of propoxyphene, is used as a reference for distinguishing a preliminary positive from negative samples. A sample with a change in absorbance ($\Delta A/\text{min}$) equal to or greater than that obtained with the cutoff calibrator is considered a preliminary positive. A sample with a change in absorbance ($\Delta A/\text{min}$) lower than that obtained with the cutoff calibrator is considered negative.

Semi-Quantitative: The semi-quantitative mode is for purposes of (1) enabling laboratories to determine an appropriate dilution of the specimen for verification by a confirmatory method such as GC/MS, LC/MS or (2) permitting laboratories to establish quality control procedures. When an approximation of concentration is required, a calibration curve can be established with five calibrators. The concentration of propoxyphene in the sample may then be estimated from the calibration curve.

Limitations

1. A preliminary positive result from the assay indicates only the presence of propoxyphene. The test is not intended for quantifying this single analyte in samples.
2. A preliminary positive result does not necessarily indicate drug abuse.
3. A negative result does not necessarily mean a person did not take illegal drugs.
4. Care should be taken when reporting results, as numerous factors (e.g., fluid intake, endogenous or exogenous interferents) may influence urine test results.
5. Preliminary positive results should be confirmed by other affirmative, analytical chemistry methods (e.g., chromatography), preferably GC/MS or LC/MS.
6. The test is designed for use with human urine only.
7. The test is not for therapeutic drug monitoring.

Typical Performance Characteristics

The results shown below were performed with a single Hitachi 717 automated clinical chemistry analyzer.

Precision:

Qualitative analysis: The three calibrators and two levels of controls were evaluated. Typical results ($\Delta A/\text{min}$) are as follows:

Concentration	Within Run (N=21)			Run-to-Run* (N=12)		
	Mean	SD	% CV	Mean	SD	% CV
0 ng/mL	117.4	0.5	0.5 %	116.8	1.0	0.9 %
225 ng/mL	225.1	1.3	0.6 %	220.8	2.4	1.1 %
300 ng/mL	261.3	1.6	0.6 %	255.9	2.1	0.8 %
375 ng/mL	287.7	1.5	0.5 %	285.1	2.2	0.8 %
1000 ng/mL	350.0	1.4	0.4 %	349.5	1.9	0.6 %

*Run-to-Run testing completed over 3 weeks

Semi-quantitative analysis: The concentrations of the cutoff level and the two levels of controls were determined with reference curves from five calibrators. Typical results (ng/mL) are as follows:

Concentration	Within Run (N=21)			Run-to-Run* (N=12)		
	Mean	SD	% CV	Mean	SD	% CV
225 ng/mL	231.3	3.1	1.3 %	232.6	3.0	1.3 %
300 ng/mL	299.6	5.8	1.9 %	298.7	4.7	1.6 %
375 ng/mL	379.7	5.6	1.5 %	378.0	7.4	2.0 %

*Run-to-Run testing completed over 3 weeks

Sensitivity: Sensitivity, defined as the lowest concentration that can be differentiated from negative urine with 95 % confidence, was tested to be 7.5 ng/mL.

Accuracy: One hundred and forty-six (146) clinical urine specimens were tested with the LZI Propoxyphene EIA. Forty-eight (48) samples were found positive and 98 samples were found negative. All positive samples were confirmed with GC/MS.

Cutoff Value (300 ng/mL)	GC/MS	LZI PPX EIA	% Agreement with Predicate
# Positive Samples	48	48	100 %
# Negative Samples	98	98	100 %
Total # of Samples	146	146	N/A

In addition to the above study, 18 samples were diluted to further characterize performance. The diluted samples contained propoxyphene concentrations ranging from 68 ng/mL to 550 ng/mL and norpropoxyphene concentrations ranging from 722 ng/mL to > 10,000 ng/mL. All 18 samples were tested positive in the EIA.

Further dilutions were performed with seven of the 18 samples to give concentrations of propoxyphene ranging from 17 ng/mL to 50 ng/mL and norpropoxyphene ranging from 167 ng/mL to 372 ng/mL. All seven samples tested negative with the current EIA.

Analytical Recovery: In qualitative analysis, the assay correctly identified spiked samples containing more than 300 ng/mL of propoxyphene (n=25, spiked levels equal to or higher than Level 2 Control) as positive, and those containing less than 300 ng/mL of propoxyphene (n=25, spiked levels equal to or less than Level 1 Control) as negative.

For semi-quantitative analysis, the average recovery for samples spiked with 30 ng/mL to 900 ng/mL (five samples at each level) of propoxyphene is summarized in the following table:

Expected Value (ng/mL)	Observed Value (ng/mL)	% Recovery
30	31.9	106.5 %
60	62.3	103.8 %
120	127.7	106.4 %
180	188.2	104.6 %
225	236.0	104.9 %
375	389.2	103.8 %
500	506.6	101.3 %
600	602.2	100.4 %
750	719.3	95.9 %
900	823.1	91.5 %

Specificity: Various potentially interfering substances were tested for cross-reactivity with the assay. Test compounds were spiked into the drug-free urine calibrator matrix to various concentrations and evaluated against the cutoff calibrator. The table below lists the concentration of each test compound that gave a response approximately equivalent to that of the cutoff calibrator (as positive) or the maximal concentration of the compound tested that gave a response below the response of the cutoff calibrator (as negative).

Structurally Related Propoxyphene Compounds:

Compound	Target [\square] (ng/mL)	% Cross-Reactivity
Propoxyphene	300	Positive
Norpropoxyphene	620	Positive

Structurally Unrelated Pharmacological Compounds:

Compound	Target [\square] ($\mu\text{g/mL}$)	% Cross-Reactivity
Acetaminophen	1000	Negative
Acetylsalicylic Acid	1000	Negative
Amitriptyline	1000	Negative
Amobarbital	1000	Negative
Amphetamine	1000	Negative
Benzoyllecgonine	1000	Negative
Bupropion	1000	Negative
Caffeine	1000	Negative
Chlorpheniramine	1000	Negative
Chlorpromazine	1000	Negative
Cocaine	1000	Negative
Codeine	1000	Negative
Dextromethorphan	1000	Negative
Ecgonine	1000	Negative
Ephedrine	1000	Negative
Imipramine	1000	Negative
Lidocaine	1000	Negative
Meperidine	1000	Negative
Methadone	1000	Negative
Methamphetamine	1000	Negative
Methaqualone	1000	Negative
Morphine	1000	Negative
Nortriptyline	1000	Negative
Oxazepam	1000	Negative
Phencyclidine	1000	Negative
Phenobarbital	1000	Negative
Promethazine	1000	Negative
Ranitidine	1000	Negative
Secobarbital	1000	Negative
Valproic Acid	1000	Negative

It is possible that other substances and/or factors not listed above may interfere with the test and cause false positive results.

Bibliography:

1. Urine Testing for Drug of Abuse, National Institute on Drug Abuse (NIDA) Research Monograph 73, (1986).
2. Mandatory Guidelines for Federal Workplace Drug Testing Program, National Institute on Drug Abuse, Federal Register, 53(69): 11970 (1988).
3. <http://www.fda.gov/downloads/Drugs/DrugSafety/PostmarketDrugSafety/InformationforPatientsandProviders/UCM234349.pdf>
4. Goodman, L., and Gilman, A., The Pharmacological Basis of Therapeutics, 8th ed., N.Y., Pergamon Press, (1991).
5. Amsterdam, E., Rendig, S., Henderson, G., and Mason, D., Depression of Myocardial Contractile Function by Propoxyphene and Norpropoxyphene, *J. Cardiovasc. Pharm.*, 3:129-138, (1981).

Bibliography, continued:

6. Bogartz, L., and Miller, W., Pulmonary Edema Associated with Propoxyphene Intoxication, *J Am. Med. Asso.*, **215**:259-262 (1971).
7. Druid, H., and Holmgren, P., A compilation of Fatal and Control Concentrations of Drugs in Postmortem Femoral Blood, *J. For. Sci.* **42**:79-87 (1997).
8. McMahon, R., Sullivan, H., Due, S., and Marshall, F., The Metabolism Pattern of d-Propoxyphene in Man. The Use of Heavy isotopes in Drug Disposition Studies, *Life Sci.*, **12**:463-473 (1973).
9. McMahon, R., Ridolfo, A., Cupl, H., et al., The Fate of Radiocarbon-Labeled Propoxyphene in Rat, Dog, and Human, *Tox. Appl. Pharm.*, **19**:427-444 (1971).
10. Gram, L., Schou, J., Way, W., et al., d-Propoxyphene Kinetics After Single Oral and Intravenous Doses in Man, *Clin. Pharm. Ther.*, **26**:473-482 (1979).
11. Rubenstein, K.E., Schneider, R.S., and Ullman, E.F., Homogeneous Enzyme Immunoassay: A New Immunochemical Technique, *Biochem Biophys Res Commun*, **47**:846 (1972).
12. Sodium Azide. National Institute for Occupational Safety (NIOSH). Pocket Guide to Chemical Hazards. Third Printing, September 2007. Available online at: <https://www.cdc.gov/niosh/npg/default.html>
13. Yahya, A.M., McElnay, J.C., and D'Arcy, P.F., Drug absorption to glass and plastics, *Drug Metabol Drug Interact*, **6**(1):1-45 (1988).
14. Rockerbie, R.A., and Campbell, D.J., Effect of Specimen Storage and Preservation on Toxicological Analyses of Urine, *Clin. Biochem*, **11**(3): 77-81 (1978).
15. Gonzales, E., et al., Stability of pain-related medications, metabolites, and illicit substances in urine, *Clinica Chimica Acta*, **416**:80-85 (2013).
16. Nichols, J., Instrument Validation: The Road to Success. CLN's Lab 2004: From Basic to Advanced Series. 14-16 (2004).
17. CDRH Guidance for Industry and FDA Staff: Replacement Reagent and Instrument Family Policy (2003).

Additions, deletions, or changes are indicated by a change bar in the margin.
For technical assistance please call: (408) 970-8811

Manufacturer:

Lin-Zhi International, Inc.

2945 Oakmead Village Court

Santa Clara, CA 95051

USA

Tel: (408) 970-8811

Fax: (408) 970-9030

www.lin-zhi.com



Authorized European

Rep. within the EU:

CEpartner4U

Esdoornlaan 13

3951DB Maarn

The Netherlands

www.cepartner4u.eu



© June 2019 Rev. 8

Printed in USA