

# LZI Negative Calibrator



IVD For In Vitro Diagnostic Use Only



## Lin-Zhi International, Inc.

REF	Description	Quantity
0001	Negative Calibrator	1 x 5 mL

### Intended Use

The Lin-Zhi International (LZI) Negative Calibrator is for use as the negative calibrator in the qualitative and semi-quantitative calibration of a number of LZI Enzyme Immunoassays on a number of automated clinical chemistry analyzers (1).

### Description of the Calibrator:

The LZI Negative Calibrator is a human urine-based ready to use liquid. It is a processed drug-free human urine matrix containing buffers, stabilizers, and less than 0.1 % of sodium azide (2).

### Precautions and Warning

- *The LZI Negative Calibrator is for in vitro diagnostic use only. Harmful if swallowed.*
- *The calibrators contain sodium azide, which may react with lead or copper plumbing to form potentially explosive metal azide. When disposing such liquids always flush with a large volume of water to prevent azide build-up.*
- *The calibrator is prepared from non-sterile human urine. It is not tested by licensed reagents for the presence of antibodies to human immunodeficiency viruses, the hepatitis antigens, and/or anti-hepatitis antibodies. It should be handled as potentially infectious. Always follow good laboratory practice to avoid any skin contact or ingestion.*
- *Do not use the calibrator beyond its expiration date.*

### Preparation and Storage

The calibrator is provided ready-to-use. No reconstitution is required. Label the cap before removal to identify it with the original bottle. The calibrator should be stored refrigerated at 2-8°C when not in use.

### Stability

When stored refrigerated at 2-8°C, the calibrator is stable either opened-recapped or unopened until the expiration date printed on the bottle label. Store calibrator tightly capped when not in use. Calibrator solution dispensed in the sample cups and left on board of the clinical analyzer should be discarded after use.

### Procedure and Results

Recalibration should be performed after reagent bottle change or if there is a change in calibrator or reagent lot, and after instrument maintenance is performed. For interpretation of results, refer to the appropriate LZI Enzyme Immunoassay package insert.

### Limitations

The LZI Negative Calibrator is for use with a number of LZI Enzyme Immunoassays.

### Bibliography

1. Urine testing for Drug of Abuse. National Institute on Drug Abuse (NIDA) Research Monograph 73, 1986.
2. Sodium Azide. National Institute for Occupational Safety (NIOSH) Pocket Guide to Chemical Hazards. Available online at: <https://www.cdc.gov/niosh/npg/npgd0560.html>

**Notice: Adulteration of reagents, use of instruments without appropriate capabilities, or other failure to follow instructions as set forth in this labeling can affect performance characteristics, and stated or implied label claims.**



**Manufacturer:**  
**Lin-Zhi International, Inc.**  
2945 Oakmead Village Court  
Santa Clara, CA 95051  
USA  
Tel: (408) 970-8811  
Fax: (408) 970-9030  
[www.lin-zhi.com](http://www.lin-zhi.com)

### EC REP Authorized European Rep. within the EU:

CEpartner4U  
Esdoornlaan 13  
3951 DB Maarn  
The Netherlands  
[www.cepartner4u.eu](http://www.cepartner4u.eu)



© April 2018 Rev. 0

Printed in USA